

**EC DECLARATION OF CONFORMITY**

*Date of Issuance: January 10, 2023*

*Valid till: January 09, 2027*

**SU Instruments (PVT) Ltd.**

Captain Shukar-Din Street, Kotli Behram Stop, Kashmir Road,  
Sialkot 51310, Pakistan.

Tel: +92-316-7988228, Fax: +92-52-4266356

Email: [info@su-instruments.com](mailto:info@su-instruments.com)

Web: [www.su-instruments.com](http://www.su-instruments.com)

Is registered with EUDAMED with SRN number **PK-MF-000035607**,

and competent authority **“DE/CA73 - Landesamt für Gesundheit und Soziales Berlin, Germany”**

Hereby declares that EU Representative had been appointed as follow for Medical Devices (Class I- Reusable)

**YELLOW’S MEDICAL INSTRUMENTS GMBH**

Max-Planck Str. 3, D-12489 Berlin,

Germany

Tel: +49-30 234 58 995

Fax: +49-30 288 70 554

E.mail: [info@cemark-y-medical.com](mailto:info@cemark-y-medical.com)

web: [www.cemark-y-medical.com](http://www.cemark-y-medical.com)

We hereby declare that EC Declaration of Conformity is issued under the sole responsibility of the SU Instruments as per Article 19 and Annex IV and tends to meet the essential requirements as per Annex IV under Regulation (EU) 2017/745 (MDR 2017/745 for CE). Medical Devices attached in Annex A are in conformity with MDR 2017/745 and relevant harmonized standards and the relevant parts of applicable standards of Official Journal of the European Union, are applied where applicable and presumed to be in conformity with the requirements of **“Risk Class I”** of the device in accordance with rules set out in Annex VIII covered by those standards of parts thereof.

The Products manufactured are provided in Annex A of this EC Declaration of Conformity. The products declared in Annex A are in compliance to applicable standards and or are harmonized by EU Medical Devices Regulation 2017/745 and DE/CA73 - Landesamt für Gesundheit und Soziales Berlin, Germany. The traceability of the device covered by the EU Declaration of Conformity are in compliance where appropriated (Quality Management System and UDI) protocols, as well as its intended purpose.

We also declare; Products mentioned in Annex A of this declaration are in conformity with applicable general safety and performance requirements given in Annex I of MDR 2017/745 under the conditions of the intended use of the device. Documents will be presented upon request by DE/CA73 - Landesamt für Gesundheit und Soziales Berlin, Germany, the indicated therein SU Instruments will provided that technical documentation in its entirety and summary thereof.

The products manufactured are Reusable-Class I (Non-Sterilized), low risk products in line with Annex VIII;

The products referred to Annex A are developed with due care in lieu of technical documentation referred to in Annex II and III of MDR 2017/745 and harmonized standards given in Annex B of this declaration.

To keep the technical documentation, the EU declaration of conformity and relevant certificates, including any amendments and supplements, issued in accordance with the article 56, available for the DE/CA73 - Landesamt für Gesundheit und Soziales Berlin, Germany for period of at least 10 years after the last device covered by the EU declaration of conformity has been placed on the market.

The Basic UDI-DI as referred to in Part C of Annex VI had been established in reference to the products in Annex A as per information to be submitted upon the registration of Devices and Economic Operators in accordance with article 28, 29 and 31.



SU. Instruments (PVT)  
CEO

Captain Shukar-Din Street, Kotli Behram Stop, Kashmir Road, Sialkot 51310, Pakistan.

Tel: 0092-52-4266356 Email: [info@su-instruments.com](mailto:info@su-instruments.com) Web: [www.su-instruments.com](http://www.su-instruments.com)

Annex A

Products List

Declares that products list has been classified as Medical Devices (Class I-Reusable) inclusive of mentioned products variant s/s in conformity with the essential requirements, provisions of Medical Devices Regulation 2017/745 and DE/CA73 - Landesamt für Gesundheit und Soziales Berlin, Germany has been appointed as EC Representative for following products.

Sr. No	Product Group	Risk Class	PRODUCTS INFORMATIONN	PRODUCTS CODE (SUI)	PRODUCTS CODE (CUSOMER)	BASIC UDI-DI
1	SPECULUM	CLASS 1	CUSCO SPECULUM LARG FIG. 3 (110 X 37 MM)	SUEEL0444	VS-200-03	4055447002342

Annex B

*The products mentioned in Annex A of this declaration are hereby declared in conformity with applicable harmonized standards mentioned below:*

Standards	Description
Medical devices-Application of risk management to medical devices	ISO 14971
Medical Device Regulation (MDR)	Regulation (EU) 2017/745
Medical Devices quality management systems-Requirements for regulatory Purpose.	EN ISO 13485: 2016
Biological evaluation of medical devices-Part 1: Evaluation and testing	ISO 10993-1
Biological evaluation of medical devices-Part 2: Animal welfare requirements	ISO 10993-2
Biological evaluation of medical devices-Part 3: Test for genotoxicity, carcinogenicity and reproductive toxicity	ISO 10993-3
Medical Devices-- symbols to be used with medical device labels, labeling, and information to be supplied – Part2: Symbols	ISO 15223-22:2010
Graphical Symbols for use in the labeling of medical devices	ISO 15223-1
Guide line for authorized representative	MEDDEV 2.5./10
Medical Devices classification	MEDDEV 2.4/1
Evaluation of clinical data-guide for manufacturer and notified bodies	MEDDEV 2.7.1 Appendix-1
Medical device vigilance system	MEDDEV 2.12.1
Standard specification for stainless steel billet, bar, and wire for Surgical Instruments	ASTM-F899
Surgical Instruments—Metallic Materials—Part 1: Stainless Steel	ISO 7153-1
Sterilization of Health care products	ISO 17665-1
Competent Authority (DE/CA73 - Landesamt für Gesundheit und Soziales Berli	



Sd. Instruments (Pvt)  
CEO